

---

# TGFBR2-Guided Biomaterial Scaffolds for Aortic Root Reconstruction in Loeys-Dietz Syndrome: Toward Precision Aortopathy Repair

---

Muhammad Moin Ud Din Arshad <sup>1,2</sup>, Maryam Ahmad Khan <sup>3</sup>, Sana Tariq <sup>1</sup>, Abaidullah Shaukat <sup>1</sup>, Ezza Ikram <sup>1</sup>

<sup>1</sup> Quaid-e-Azam Medical College, Bahawalpur, Pakistan

<sup>3</sup> Shahida Islam Medical College, Lodhran, Pakistan

<sup>2</sup>Corresponding Author: [Moinali596@gmail.com](mailto:Moinali596@gmail.com)

## ABSTRACT:

Loeys–Dietz syndrome (LDS) is a genetically mediated aortopathy caused by pathogenic variants in TGF- $\beta$  pathway genes (including TGFBR1 and TGFBR2) and is characterized by aggressive aortic root dilation, dissection at young ages, and high reoperation rates. Current surgical conduits restore gross structural integrity but remain biologically inert, failing to correct the cell-level signaling abnormalities that drive downstream disease. We propose a translational framework for TGFBR2-guided biomaterial scaffolds that combines scaffold mechanics with locally targeted molecular modulation of TGF- $\beta$ /SMAD signaling to promote adaptive behavior of smooth muscle cells (SMCs) and extracellular matrix (ECM) homeostasis. This Perspective defines specific molecular candidates (ligand-mimetic peptides, receptor-biasing aptamers/antibodies, and miRNA payloads such as miR-145/miR-21/miR-29 modulation), delivery and tethering strategies, provisional release kinetics, and safety/mitigation measures. We specify prespecified mechanical and biological acceptance criteria (biaxial testing, pulsatile loop validation, compliance, suture retention, burst pressure, MMP assays, pSMAD2/3 monitoring), computational modeling approaches (Holzapfel–Gasser–Ogden constitutive model, fluid–structure interaction, ML surrogates), and a staged preclinical and regulatory pathway including GLP and cGMP considerations for combination products. Clear predefinition of endpoints, power assumptions, and regulatory engagement are essential to advance these scaffolds from concept to Early Feasibility Study (EFS) and human translation in high-risk LDS patients.

**KEYWORDS:** *Loeys–Dietz syndrome; TGFBR2; biomaterial scaffold; aortic root; tissue engineering; combination product*

## CLINICAL NEED:

Loeys–Dietz syndrome (LDS) encompasses a spectrum of aggressive thoracic aortopathies driven by mutations in genes of the TGF- $\beta$  signaling axis

(notably TGFBR1 and TGFBR2), producing medial degeneration, elastin fragmentation, and maladaptive remodeling of the aortic root. Despite advances in surgical techniques and graft design, available implants (Dacron Valsalva conduits,

composite valve grafts, decellularized homografts, and ePTFE) are mechanically effective but biologically passive. In genetically susceptible patients, this biologic inertness contributes to ongoing remodeling of the residual aorta and necessitates lifetime surveillance and repeated interventions (1,3). A targeted scaffold that restores both structure and key signaling homeostasis could reduce downstream dilation and the need for reoperation.

### Design Rationale:

“TGFBR2-guided” scaffolds intentionally present microenvironmental cues that bias local TGF- $\beta$  receptor signaling toward adaptive, homeostatic responses in aortic SMCs. Specific and named candidate strategies include:

- **Ligand-mimetic peptides** that reproduce receptor interaction motifs or recruit co-receptors, presented by covalent tethering (click chemistry; EDC/NHS) or affinity immobilization (heparin/gelatin domains) to allow controlled desorption and spatially constrained activity.
- **Aptamers or receptor-biasing antibodies** localized to the scaffold surface to sterically modulate TGFBR2 interactions (designed to bias complex formation rather than wholesale blockade).
- **miRNA-based modulation:** deliver miR-145 mimics to support contractile SMC phenotype, inhibit miR-29 family members to restore elastogenic programs, and carefully titrate miR-21 modulation given its context dependence; all payload choices should be directed by expression profiling of patient-derived or genetically representative SMCs (4–6).

### Delivery Platforms:

Electrospun fiber cores embedding PLGA or lipid nanoparticles (LNPs), layer-by-layer polyelectrolyte films, hydrogel interlayers, or nanoporous coatings.

### Working Release Target:

Initial burst <20% in first 48–72 h, sustained therapeutically active release for ~2–8 weeks (window of early remodeling); final kinetics to be tuned by in vitro PK/PD and ex vivo bioreactor data (7).

**Table 1:** Design levers, biological targets, mechanical criteria, and acceptance standards

Design Lever	Biological Target	Mechanical Target	Pass / Acceptance Criterion	Test Method
Ligand-mimetic peptide tethering (covalent)	Local modulation of TGFBR2 signaling → normalize pSMAD2/3 & SMC phenotype	Compliance matched to native aortic root within $\pm 20\%$	pSMAD2/3 within $\pm 25\%$ of control/healthy explant; compliance within $\pm 20\%$ of native	IHC/Western for pSMAD2/3; biaxial mechanical testing (stress-strain)
Aptamer/receptor-biasing antibody presentation	Reduce maladaptive ligand activity without systemic blockade	Suture retention $\geq$ native tissue baseline (no clinically meaningful handling deficit)	No increase in neointima score vs comparator; suture retention $\geq$ native baseline	Histology (EVG), suture pull-out testing (N)
miRNA payload (e.g., miR-145 mimic; miR-29 inhibition)	Promote contractile SMC phenotype and elastogenesis; reduce MMP activity	Minimal change in burst pressure vs comparator; fatigue $\geq 10^7$ cycles	$\geq 30\%$ reduction in pathological MMP activity (MMP-2/9) vs ePTFE at prespecified timepoint; fatigue passes $10^7$ cycles without rupture	Zymography/ELISA for MMPs; dynamic fatigue bench test
Fiber orientation / porosity tuning	Favor in-growth and alignment of SMCs; limit pathological shear	Local strain hotspots < defined threshold (from FEA)	No localized overstretch > pre-defined strain; recellularization score $\geq$ comparator	FEA (HGO model) predictions; histologic cell density scoring
Dual-layer scaffold (structural + bioactive)	Fail-safe mechanical support if bioactivity attenuates	Outer layer maintains integrity if inner layer bioactivity lost	Structural integrity maintained through simulated worst-case bioactive loss	Pulsatile loop rupture/creep testing; long-term soak testing

## Safety, Risk Mitigation, and Biomarkers:

Because TGF- $\beta$  signaling is contextually pleiotropic, both over- and under-modulation present risks (fibrosis, intimal hyperplasia, impaired elastogenesis), a built-in risk-mitigation strategy must include:

1. **Predefined biomarkers:** pSMAD2/3 by immunohistochemistry/Western blot, SMC markers (ACTA2, MYH11, TAGLN), elastin/collagen ratio by biochemical assays and EVG, MMP-2/9 activity (zymography/ELISA), macrophage infiltration (CD68), thrombosis and calcification scoring. Time points should be prespecified (e.g., 7, 14, 28, 90 days) with criteria for escalation or study halt (2).
2. **Engineering controls include** formulation-defined release ceilings, reversible/affinity coatings that permit passive attenuation, and dual-layer scaffolds where an outer structural layer ensures mechanical continuity regardless of bioactive performance.
3. **Operational rules:** in vitro dose finding, conservative starting doses in vivo, and prespecified stopping rules tied to biomarker excursions (for example,  $>2\times$  baseline pSMAD2/3 or pathological neointimal thickness beyond predefined limits).

These safeguards should be documented in study protocols and regulatory briefing packages.

## Comparator Selection and Mechanical Acceptance Criteria:

Comparators must reflect current clinical standards: ePTFE, woven Dacron (including Valsalva designs), and decellularized homografts. Mechanical/ex vivo acceptance metrics should be prespecified, including compliance ( $\text{mmHg}^{-1}$ ) aligned with native root values, suture retention strength (N), burst pressure (mmHg), fatigue resistance to  $\geq 10^7$  cycles, and creep under physiologic loads. Given the aortic root's non-cylindrical geometry, testing must include biaxial mechanical testing, coronary-button

reimplantation simulation, and pulsatile loop validation reproducing sinus flow and valve-scaffold interactions before animal studies. Constitutive modeling should employ an anisotropic fiber-reinforced model such as Holzapfel–Gasser–Ogden (HGO) to inform fiber orientation, porosity, and compliance matching (8,7).

## Computational Workflows: FEA and ML:

Finite element analysis (FEA) must use validated constitutive descriptions (HGO) with physiologic boundary conditions (root, sinotubular junction compliance, coronary ostia constraints) and, where feasible, fluid–structure interaction to predict leaflet motion and wall shear stresses. Machine learning (ML) can accelerate optimization: train surrogates on DoE outputs where labels include mechanical failure modes (suture pullout, localized overstretch), predicted strain hotspots, and in vitro release kinetics under pulsatile conditions. Iteration cycles should be gated by simulation-derived risk scores and verified experimentally in a hierarchy (in vitro  $\rightarrow$  ex vivo  $\rightarrow$  in vivo).

## Preclinical Testing Plan and Sample-size Assumptions:

**In vitro:** experiments on patient-derived LDS SMCs and control SMCs with endpoints at 7/14/28 days: ACTA2/MYH11/TAGLN expression, elastin and collagen gene/protein levels, MMP-2/9 activity, viability, and proliferation assays. Use dose–response and transcriptomic profiling to select payload and dose.

**Ex vivo:** donor or porcine aortic ring testing in bioreactors to confirm compliance, suture behavior, and release under pulsatile flow.

**Small animal:** genetically accurate Tgfb $\beta$ 2 murine models (conditional or pathogenic missense variants) for recellularization, pSMAD2/3, elastin architecture, and early safety signals.

**Large animal:** porcine aortic root replacement with coronary button reimplantation to test surgical handling, hemodynamics, and durability.

**Sample-size Example:** To detect a 20% difference in diameter change at 6 months (SD = 15%,  $\alpha = 0.05$ , power = 0.8) requires  $\approx$ approximately nine (9) animals/group (round up to 10). Histologic endpoints with Cohen's  $d \approx$  approximately 1.0 may require  $\approx$ approximately 16/group in small animal models; pilot data should refine these numbers, and all analyses must be prespecified and controlled for multiplicity.

**Table 2:** Preclinical endpoints, assays, timepoints, pass/fail criteria, and planned sample size

Preclinical Stage	Endpoint (readout)	Assay / Measure	Timepoint(s)	Pass / Fail Criterion	Planned (n)
In vitro (discovery)	SMC contractile phenotype	qPCR / Western for ACTA2, MYH11, TAGLN	7, 14, 28 days	$\geq 1.5$ -fold increase in contractile markers vs untreated LDS SMCs	n = 3–6 biological replicates per condition
In vitro (discovery)	MMP activity	Zymography and ELISA (MMP-2, MMP-9)	7, 14, 28 days	$\geq 30\%$ reduction vs ePTFE control	n = 3–6
Ex vivo (engineering)	Compliance and suture performance	Biaxial testing; suture pull-out	Single session; under pulsatile flow	Compliance within $\pm 20\%$ native; suture retention $\geq$ native baseline	n = 3–5 samples per design
Small animal (proof-of-concept)	pSMAD2/3 & ECM architecture	IHC (pSMAD 2/3), EVG staining, MMP assays	4, 8, 12 weeks	No pathological increase in pSMAD 2/3 vs sham; improved elastin organization score	n $\approx$ 10–16 per group
Large animal (GLP)	Root diameter stability & surgical handling	Imaging (echo/CT), surgical assessment	1, 3, 6, 12 months	Diameter change $\leq$ comparator threshold; no technique-limiting handling issues	n $\approx$ 8–12 per group
Large animal (GLP)	Durability & thrombosis/calcification	Explant histology, gross exam, blood markers	6, 12 months	No uncontrolled thrombosis; calcification score $\leq$ comparator	n $\approx$ 8–12

*Values and sample sizes are indicative and will be refined following pilot validation.*

### Regulatory, Manufacturing, and Testing Pathway:

When a scaffold delivers biologics or oligonucleotides, it will likely be regulated as a combination product; early engagement with the appropriate regulatory agency (Pre-IDE/Q-Submission in the U.S.) is essential to determine the lead center and data expectations. Beyond ISO 10993-1, blood-contacting implants require a comprehensive suite of hemocompatibility tests, including cytotoxicity, sensitization, irritation, hemolysis, pyrogenicity, systemic toxicity, genotoxicity, implantation, and extractables/leachables evaluation. Device mechanical standards (burst, fatigue, compliance) and validated sterilization methods (EtO or gamma irradiation with functional retention of ligands/miRNA) must be defined, accompanied by lot-release assays (dose uniformity, endotoxin, and functional bioactivity by in vitro assay). GLP animal studies and cGMP processes for the biologic component are prerequisites for EFS/IDE submissions (9–11).

### Combination Product and GMP Implications:

The inclusion of miRNA or biologic payloads requires cGMP for the drug/biologic component and validated aseptic finishing for the final combination product. Early classification (device, drug, or biologic) determines which regulatory office leads (CDRH, CDER, or CBER in the U.S.) and frames the preclinical and clinical data package; formal interaction with regulators should be sought early. Stability testing (both real-time and accelerated),

validation of cold chain requirements as needed, and demonstration of dose uniformity are essential elements of the manufacturing control strategy.

### **Clinical Entry and First-in-human Considerations:**

First-in-human translation should proceed under an Early Feasibility Study (EFS)/IDE framework—not primary compassionate use—with a narrowly defined high-risk LDS cohort lacking standard alternatives. The primary aims should be safety and feasibility, with short-term imaging and biomarker endpoints (6–12 months). Prespecified stopping rules (based on pSMAD2/3 elevations, pathological neointima, or thrombosis) and transparent consent, emphasizing investigational risks and the potential need for reoperation, are essential.

### **Measurable Endpoints and Statistical Plans (examples):**

**In vitro:** fold change in SMC contractile markers (ACTA2/MYH11/TAGLN), elastin/collagen ratio, MMP-2/9 activity at 7/14/28 days.

**Ex vivo/in silico:** compliance ( $\text{mmHg}^{-1}$ ), suture retention (N), burst pressure (mmHg), fatigue to  $\geq 10^7$  cycles.

**In vivo:** root diameter stability by imaging, elastic lamellae integrity (EVG), pSMAD2/3 quantification, CD68 inflammatory index, thrombosis/calcification scoring, and MAE/MACE rates. For claims such as “ $\geq 30\%$  reduction in pathological MMP activity,” prespecify which MMP(s), the assay (zymography densitometry or validated ELISA), time points, n, and the statistical framework (ANOVA with correction or two-tailed tests) (2).

### **Roadmap and Milestones:**

A pragmatic staged program: (1) discovery and in vitro candidate screening in patient-derived SMCs (dose–response and PK/PD); (2) ex vivo engineering validation (pulsatile loop, biaxial testing; FEA tuning); (3) small-animal proof-of-concept with

prespecified powering and GLP elements; (4) large-animal GLP studies for surgical and durability validation; (5) regulatory engagement and EFS/IDE with cGMP and lot-release documentation.

### **CONCLUSION:**

TGFBR2-guided scaffolds represent a hypothesis-driven convergence of molecular genetics and advanced biomaterials, aiming to restore signaling homeostasis while providing durable mechanical support in LDS aortopathy. To transition from concept to clinic, these programs must enumerate named biomolecular candidates and delivery strategies, predefine biomechanical and biological acceptance criteria, implement robust safety/mitigation plans, and follow GLP/cGMP regulatory pathways. With rigorous preclinical validation and early regulatory partnership, TGFBR2-guided scaffolds have the potential to impact the management of genetically mediated aortic disease significantly.

### **DECLARATIONS**

#### **Conflict of Interest:**

Not applicable; no human or animal data are presented.

#### **Ethical Considerations:**

Not applicable; planned in vivo studies will undergo IACUC/ethics committee approval prior to initiation.

#### **Declaration of AI use:**

This manuscript was drafted and revised with the assistance of an AI language model (ChatGPT, GPT-5, OpenAI) for editing, organization, and clarity. All intellectual content, interpretation and final approval are the sole responsibility of the named authors.

**Funding:** None

## REFERENCES

1. Verstraeten A, Dietz HC, Loeys BL. Loeys–dietz syndrome. Cassidy and Allanson's Management of Genetic Syndromes. 2021 Feb 19:563-76. <https://doi.org/10.1002/9781119432692.ch36>
2. Chen, J., & Chang, R. (2022). Association of TGF- $\beta$  Canonical Signaling-Related Core Genes With Aortic Aneurysms and Aortic Dissections. *Frontiers in Pharmacology*, 13, 888563. <https://doi.org/10.3389/fphar.2022.888563>
3. Yang, G., Khan, A., Liang, W., Xiong, Z., & Stegbauer, J. (2024). Aortic aneurysm: pathophysiology and therapeutic options. *MedComm*, 5(9), e703. <https://doi.org/10.1002/mco2.703>
4. Zhang C. (2009). MicroRNA-145 in vascular smooth muscle cell biology: a new therapeutic target for vascular disease. *Cell Cycle (Georgetown, Tex.)*, 8(21), 3469–3473. <https://doi.org/10.4161/cc.8.21.9837>
5. Maegdefessel, L., Azuma, J., Toh, R., Deng, A., Merk, D. R., Raiesdana, A., Leeper, N. J., Raaz, U., Schoelmerich, A. M., McConnell, M. V., Dalman, R. L., Spin, J. M., & Tsao, P. S. (2012). MicroRNA-21 blocks abdominal aortic aneurysm development and nicotine-augmented expansion. *Science Translational Medicine*, 4(122), 122ra22. <https://doi.org/10.1126/scitranslmed.3003441>
6. Maegdefessel, L., Azuma, J., & Tsao, P. S. (2014). MicroRNA-29b regulation of abdominal aortic aneurysm development. *Trends in Cardiovascular Medicine*, 24(1), 1–6. <https://doi.org/10.1016/j.tcm.2013.05.002>
7. Xiong, M., Yang, L., Liu, X., Luo, S., & Wang, Y. (2025). Circumferentially Aligned Electrospun Vascular Grafts Improves Its Vascular Regeneration and Remodeling in vivo. *International journal of nanomedicine*, 20, 7515–7532. <https://doi.org/10.2147/IJN.S518283>
8. Holzapfel GA, Gasser TC, Ogden RW. A new constitutive framework for arterial wall mechanics and a comparative study of material models. *Journal of elasticity and the physical science of solids*. 2000 Jul;61(1):1-48. <https://link.springer.com/article/10.1023/A:1010835316564>
9. U.S. Food and Drug Administration. Use of ISO 10993-1, Biological evaluation of medical devices—Part 1: Evaluation and testing within a risk management process: Guidance for industry and FDA staff. Silver Spring (MD): FDA; 2023. <https://content.govdelivery.com/accounts/USFDA/bulletins/36ec9b7>
10. U.S. Food and Drug Administration, Office of Combination Products. Guidance materials and updates. Silver Spring (MD): FDA; 2025. <https://www.fda.gov/combination-products>
11. U.S. Food and Drug Administration. Current Good Manufacturing Practice requirements for combination products. Silver Spring (MD): FDA; 2017 [updated 2024–2025]. <https://www.fda.gov/drugs/pharmaceutical-quality-resources/current-good-manufacturing-practice-cgmp-regulations>